

**TEST REPORT**

REPORT NO:-QMTL/R-85413

<b>CLIENT:- CRYSTALLINE ENVIRONMENTAL SERVICES</b> P.O.Box.131463, Abu Dhabi, U.A.E Ph:02-6508930/0557167017		
<b>ATTN:MR. JASMAN PINTO</b>		
Product: Water	Sampling Method: APHA 9060 A	Sample ID: QMTL/S-85413
Reference: <b>HEALTHPOINT HOSPITAL</b>	Sampled By: QMTL Representative	Date Received: 19.12.2019
<b>Sample Description :</b> Dental Chair- 1 Sample Onsite Temp: 24°C Sample Receipt Temp: 4°C	Source: Dental Chair	Date(s) Tested: 19-21.12.2019
	Sample Container: 1 X 500ml PB	Date Reported: 23.12.2019
	Sample Appearance: Clear	Date of Sampling/Time: 19.12.2019/11.39 AM
On-Site Treatment: NA	Tested by Emp#: 55 & 74	Location: AUH

**TEST RESULTS**

PARAMETERS	TEST METHODS (APHA 23 <sup>RD</sup> EDITION)	UNITS	SPECIFICATIONS	RESULTS
<b>MICROBIOLOGICAL ANALYSIS</b>				
Total Bacterial Count	APHA 9215 B	CFU/ml	10@37°C	<1
Total Coliforms	APHA 9222 B	CFU/100ml	Zero	<b>Zero</b>
Faecal coliform	APHA 9222 D	CFU/100ml	Zero	<b>Zero</b>
Pseudomonas aeruginosa	APHA 9213 E	CFU/100ml	Zero	<b>Zero</b>
Escherichia coli	APHA 9222 H & I	CFU/100ml	Zero	<b>Zero</b>

REMARKS : The sample meets the limits of HAAD EHSMS Standard for Healthcare Providers- RSB Water Quality Regulations (4<sup>th</sup> Edition) January 2014 requirements for potable water with respect to the above tests only.

Key Words:- CFU : Colony Forming Unit, APHA : American Public Health Association, RSB : Regulation & Supervision Bureau HAAD : The Health – Authority- Abu Dhabi, EHSMS : Environment, Health and Safety Management System, NA : Not Applicable

Report prepared by Emp#: 75

Test Method deviation : None

The above test results are only applicable to the sample(s) referred above

Reported By

**ANILA GEORGE**  
**LABORATORY MANAGER**



The test-Report shall not be reproduced (except in full) without the written approval of QMTL. When analysis is witnessed by us or carried out by sub contract labs, our responsibility is only to ensure that the analysis is conducted to standard test methods in accordance with industry accepted practice.

END OF REPORT

**TEST REPORT**

REPORT NO:-QMTL/R-85414

<b>CLIENT:- CRYSTALLINE ENVIRONMENTAL SERVICES</b> P.O.Box.131463, Abu Dhabi, U.A.E Ph:02-6508930/0557167017		
<b>ATTN:MR. JASMAN PINTO</b>		
Product: Water	Sampling Method: APHA 9060 A	Sample ID: QMTL/S-85414
Reference: <b>HEALTHPOINT HOSPITAL</b>	Sampled By: QMTL Representative	Date Received: 19.12.2019
<b>Sample Description :</b> Dental Chair- 2 Sample Onsite Temp: 24°C Sample Receipt Temp: 4°C	Source: Dental Chair	Date(s) Tested: 19-21.12.2019
	Sample Container: 1 X 500ml PB	Date Reported: 23.12.2019
	Sample Appearance: Clear	Date of Sampling/Time: 19.12.2019/11.48 AM
On-Site Treatment: NA	Tested by Emp#: 55 & 74	Location: AUH

**TEST RESULTS**

PARAMETERS	TEST METHODS (APHA 23 <sup>RD</sup> EDITION)	UNITS	SPECIFICATIONS	RESULTS
<b>MICROBIOLOGICAL ANALYSIS</b>				
Total Bacterial Count	APHA 9215 B	CFU/ml	10@37°C	<1
Total Coliforms	APHA 9222 B	CFU/100ml	Zero	<b>Zero</b>
Faecal coliform	APHA 9222 D	CFU/100ml	Zero	<b>Zero</b>
Pseudomonas aeruginosa	APHA 9213 E	CFU/100ml	Zero	<b>Zero</b>
Escherichia coli	APHA 9222 H & I	CFU/100ml	Zero	<b>Zero</b>

REMARKS : The sample meets the limits of HAAD EHSMS Standard for Healthcare Providers- RSB Water Quality Regulations (4<sup>th</sup> Edition) January 2014 requirements for potable water with respect to the above tests only.

Key Words:- CFU : Colony Forming Unit, APHA : American Public Health Association, RSB : Regulation & Supervision Bureau HAAD : The Health – Authority- Abu Dhabi, EHSMS : Environment, Health and Safety Management System, NA : Not Applicable

Report prepared by Emp#: 75

Test Method deviation : None

The above test results are only applicable to the sample(s) referred above

**Reported By**

**ANILA GEORGE**  
**LABORATORY MANAGER**



The test-Report shall not be reproduced (except in full) without the written approval of QMTL. When analysis is witnessed by us or carried out by sub contract labs, our responsibility is only to ensure that the analysis is conducted to standard test methods in accordance with industry accepted practice.

**END OF REPORT**

**TEST REPORT**

REPORT NO:-QMTL/R-85415

<b>CLIENT:- CRYSTALLINE ENVIRONMENTAL SERVICES</b> P.O.Box.131463, Abu Dhabi, U.A.E Ph:02-6508930/0557167017		
<b>ATTN:MR. JASMAN PINTO</b>		
Product: Water	Sampling Method: APHA 9060 A	Sample ID: QMTL/S-85415
Reference: <b>HEALTHPOINT HOSPITAL</b>	Sampled By: QMTL Representative	Date Received: 19.12.2019
<b>Sample Description :</b> Dental Chair- 3 Sample Onsite Temp: 23.3°C Sample Receipt Temp: 4°C	Source: Dental Chair	Date(s) Tested: 19-21.12.2019
	Sample Container: 1 X 500ml PB	Date Reported: 23.12.2019
	Sample Appearance: Clear	Date of Sampling/Time: 19.12.2019/11.56 AM
On-Site Treatment: NA	Tested by Emp#: 55 & 74	Location: AUH

**TEST RESULTS**

PARAMETERS	TEST METHODS (APHA 23 <sup>RD</sup> EDITION)	UNITS	SPECIFICATIONS	RESULTS
<b>MICROBIOLOGICAL ANALYSIS</b>				
Total Bacterial Count	APHA 9215 B	CFU/ml	10@37°C	<1
Total Coliforms	APHA 9222 B	CFU/100ml	Zero	<b>Zero</b>
Faecal coliform	APHA 9222 D	CFU/100ml	Zero	<b>Zero</b>
Pseudomonas aeruginosa	APHA 9213 E	CFU/100ml	Zero	<b>Zero</b>
Escherichia coli	APHA 9222 H & I	CFU/100ml	Zero	<b>Zero</b>

REMARKS : The sample meets the limits of HAAD EHSMS Standard for Healthcare Providers- RSB Water Quality Regulations (4<sup>th</sup> Edition) January 2014 requirements for potable water with respect to the above tests only.

Key Words:- CFU : Colony Forming Unit, APHA : American Public Health Association, RSB : Regulation & Supervision Bureau HAAD : The Health – Authority- Abu Dhabi, EHSMS : Environment, Health and Safety Management System, NA : Not Applicable

Report prepared by Emp#: 75  
Test Method deviation : None  
The above test results are only applicable to the sample(s) referred above

**Reported By**

**ANILA GEORGE**  
**LABORATORY MANAGER**



The test Report shall not be reproduced (except in full) without the written approval of QMTL. When analysis is witnessed by us or carried out by sub contract labs, our responsibility is only to ensure that the analysis is conducted to standard test methods in accordance with industry accepted practice.

**END OF REPORT**

### TEST REPORT

REPORT NO:-QMTL/R-85433

<b>CLIENT:- CRYSTALLINE ENVIRONMENTAL SERVICES</b>		
P.O.Box.131463, Abu Dhabi, U.A.E Ph:02-6508930/0557167017		
<b>ATTN:MR. JASMAN PINTO</b>		
Product: Water	Sampling Method: BS 7592	Sample ID: QMTL/S-85433
Reference: <b>HEALTHPOINT HOSPITAL</b>	Sampled By: QMTL Representative	Date Received: 19.12.2019
<b>Sample Description :</b> Dental Chair- 4 Sample Onsite Temp: 23.8°C Sample Receipt Temp: 17°C	Source: Dental Chair	Date(s) Tested: 19-29.12.2019
	Sample Container: 1 X 1L PB	Date Reported: 29.12.2019
	Sample Appearance: Clear	Date of Sampling/Time: 19.12.2019/12.02 PM
On-Site Treatment: NA	Tested by Emp#: 55 & 74	Location: AUH

### TEST RESULTS

PARAMETERS	TEST METHODS	UNITS	LIMITS	RESULTS
<b>LEGIONELLA ANALYSIS</b>				
Legionella	CCFRA 3.9.1 (5 <sup>th</sup> Edition)	cfu/litre	<1	<b>Not Detected</b>
<b>IDENTIFICATION</b>				
Legionella Species	CCFRA 3.9.1 (5 <sup>th</sup> Edition)	-	-	<b>Absent</b>
Legionella pneumophila sero group 1	CCFRA 3.9.1 (5 <sup>th</sup> Edition)	-	-	<b>Absent</b>
Legionella pneumophila sero group 2-14	CCFRA 3.9.1 (5 <sup>th</sup> Edition)	-	-	<b>Absent</b>

REMARKS : The sample meets HAAD EHSMS Standard for Healthcare Providers- RSB Water Quality Regulations (4<sup>th</sup> Edition) January 2014 requirements for potable water with respect to the above tests only.

Key Words: CFU: Colony Forming Unit, BS: British Standards, CCFRA: Campden & Chorleywood Food Research Association, RSB: Regulation & Supervision Bureau, HAAD: The Health Authority- Abu Dhabi, EHSMS: Environment, Health and Safety Management System, NA: Not Applicable

Report prepared by Emp#: 75  
Test Method deviation : None  
The above test results are only applicable to the sample(s) referred above

Reported By

**ANILA GEORGE**  
**LABORATORY MANAGER**

The test Report shall not be reproduced (except in full) without the written approval of QMTL. When analysis is witnessed by us or carried out by sub contract labs, our responsibility is only to ensure that the analysis is conducted to standard test methods in accordance with industry accepted practice.

END OF REPORT

### TEST REPORT

REPORT NO:-QMTL/R-85434

<b>CLIENT:- CRYSTALLINE ENVIRONMENTAL SERVICES</b>		
P.O.Box.131463, Abu Dhabi, U.A.E Ph:02-6508930/0557167017		
<b>ATTN:MR. JASMAN PINTO</b>		
Product: Water	Sampling Method: BS 7592	Sample ID: QMTL/S-85434
Reference: <b>HEALTHPOINT HOSPITAL</b>	Sampled By: QMTL Representative	Date Received: 19.12.2019
<b>Sample Description :</b> Dental Chair- 5 Sample Onsite Temp: 23°C Sample Receipt Temp: 17°C	Source: Dental Chair	Date(s) Tested: 19-29.12.2019
	Sample Container: 1 X 1L PB	Date Reported: 29.12.2019
	Sample Appearance: Clear	Date of Sampling/Time: 19.12.2019/12.09 PM
On-Site Treatment: NA	Tested by Emp#: 55 & 74	Location: AUH

### TEST RESULTS

PARAMETERS	TEST METHODS	UNITS	LIMITS	RESULTS
<b>LEGIONELLA ANALYSIS</b>				
Legionella	CCFRA 3.9.1 (5 <sup>th</sup> Edition)	cfu/litre	<1	<b>Not Detected</b>
<b>IDENTIFICATION</b>				
Legionella Species	CCFRA 3.9.1 (5 <sup>th</sup> Edition)	-	-	<b>Absent</b>
Legionella pneumophila sero group 1	CCFRA 3.9.1 (5 <sup>th</sup> Edition)	-	-	<b>Absent</b>
Legionella pneumophila sero group 2-14	CCFRA 3.9.1 (5 <sup>th</sup> Edition)	-	-	<b>Absent</b>

REMARKS : The sample meets HAAD EHSMS Standard for Healthcare Providers- RSB Water Quality Regulations (4<sup>th</sup> Edition) January 2014 requirements for potable water with respect to the above tests only.

Key Words: CFU: Colony Forming Unit, BS: British Standards, CCFRA: Campden & Chorleywood Food Research Association, RSB: Regulation & Supervision Bureau, HAAD: The Health Authority- Abu Dhabi, EHSMS: Environment, Health and Safety Management System, NA: Not Applicable

Report prepared by Emp#: 75

Test Method deviation : None

The above test results are only applicable to the sample(s) referred above

**Reported By**

**ANILA GEORGE**  
**LABORATORY MANAGER**

The test Report shall not be reproduced (except in full) without the written approval of QMTL. When analysis is witnessed by us or carried out by sub contract labs, our responsibility is only to ensure that the analysis is conducted to standard test methods in accordance with industry accepted practice.

END OF REPORT

### TEST REPORT

REPORT NO:-QMTL/R-85435

<b>CLIENT:- CRYSTALLINE ENVIRONMENTAL SERVICES</b>		
P.O.Box.131463, Abu Dhabi, U.A.E Ph:02-6508930/0557167017		
<b>ATTN:MR. JASMAN PINTO</b>		
Product: Water	Sampling Method: BS 7592	Sample ID: QMTL/S-85435
Reference: <b>HEALTHPOINT HOSPITAL</b>	Sampled By: QMTL Representative	Date Received: 19.12.2019
<b>Sample Description :</b> Dental Chair- 6 Sample Onsite Temp: 24°C Sample Receipt Temp: 17°C	Source: Dental Chair	Date(s) Tested: 19-29.12.2019
	Sample Container: 1 X 1L PB	Date Reported: 29.12.2019
	Sample Appearance: Clear	Date of Sampling/Time: 19.12.2019/12.17 PM
On-Site Treatment: NA	Tested by Emp#: 55 & 74	Location: AUH

### TEST RESULTS

PARAMETERS	TEST METHODS	UNITS	LIMITS	RESULTS
<b>LEGIONELLA ANALYSIS</b>				
Legionella	CCFRA 3.9.1 (5 <sup>th</sup> Edition)	cfu/litre	<1	<b>Not Detected</b>
<b>IDENTIFICATION</b>				
Legionella Species	CCFRA 3.9.1 (5 <sup>th</sup> Edition)	-	-	<b>Absent</b>
Legionella pneumophila sero group 1	CCFRA 3.9.1 (5 <sup>th</sup> Edition)	-	-	<b>Absent</b>
Legionella pneumophila sero group 2-14	CCFRA 3.9.1 (5 <sup>th</sup> Edition)	-	-	<b>Absent</b>

REMARKS : The sample meets HAAD EHSMS Standard for Healthcare Providers- RSB Water Quality Regulations (4<sup>th</sup> Edition) January 2014 requirements for potable water with respect to the above tests only.

Key Words: CFU: Colony Forming Unit, BS: British Standards, CCFRA: Campden & Chorleywood Food Research Association, RSB: Regulation & Supervision Bureau, HAAD: The Health Authority- Abu Dhabi, EHSMS: Environment, Health and Safety Management System, NA: Not Applicable

Report prepared by Emp#: 75

Test Method deviation : None

The above test results are only applicable to the sample(s) referred above

**Reported By**

**ANILA GEORGE**  
**LABORATORY MANAGER**



The test Report shall not be reproduced (except in full) without the written approval of QMTL. When analysis is witnessed by us or carried out by sub contract labs, our responsibility is only to ensure that the analysis is conducted to standard test methods in accordance with industry accepted practice.

END OF REPORT

**TEST REPORT**

REPORT NO:-QMTL/R-85416

<b>CLIENT:- CRYSTALLINE ENVIRONMENTAL SERVICES</b> P.O.Box.131463, Abu Dhabi, U.A.E Ph:02-6508930/0557167017		
<b>ATTN:MR. JASMAN PINTO</b>		
Product: Water	Sampling Method: APHA 9060 A	Sample ID: QMTL/S-85416
Reference: <b>HEALTHPOINT HOSPITAL</b>	Sampled By: QMTL Representative	Date Received: 19.12.2019
<b>Sample Description :</b> Dental Chair- 7 Sample Onsite Temp: 25°C Sample Receipt Temp: 4°C	Source: Dental Chair	Date(s) Tested: 19-21.12.2019
	Sample Container: 1 X 500ml PB	Date Reported: 23.12.2019
	Sample Appearance: Clear	Date of Sampling/Time: 19.12.2019/12.12 PM
On-Site Treatment: NA	Tested by Emp#: 55 & 74	Location: AUH

**TEST RESULTS**

PARAMETERS	TEST METHODS (APHA 23 <sup>RD</sup> EDITION)	UNITS	SPECIFICATIONS	RESULTS
<b>MICROBIOLOGICAL ANALYSIS</b>				
Total Bacterial Count	APHA 9215 B	CFU/ml	10@37°C	<1
Total Coliforms	APHA 9222 B	CFU/100ml	Zero	<b>Zero</b>
Faecal coliform	APHA 9222 D	CFU/100ml	Zero	<b>Zero</b>
Pseudomonas aeruginosa	APHA 9213 E	CFU/100ml	Zero	<b>Zero</b>
Escherichia coli	APHA 9222 H & I	CFU/100ml	Zero	<b>Zero</b>

REMARKS : The sample meets the limits of HAAD EHSMS Standard for Healthcare Providers- RSB Water Quality Regulations (4<sup>th</sup> Edition) January 2014 requirements for potable water with respect to the above tests only.

Key Words:- CFU : Colony Forming Unit, APHA : American Public Health Association, RSB : Regulation & Supervision Bureau HAAD : The Health – Authority- Abu Dhabi, EHSMS : Environment, Health and Safety Management System, NA : Not Applicable

Report prepared by Emp#: 75

Test Method deviation : None

The above test results are only applicable to the sample(s) referred above

**Reported By**

**ANILA GEORGE**  
**LABORATORY MANAGER**



The test Report shall not be reproduced (except in full) without the written approval of QMTL. When analysis is witnessed by us or carried out by sub contract labs, our responsibility is only to ensure that the analysis is conducted to standard test methods in accordance with industry accepted practice.

**END OF REPORT**

**TEST REPORT**

REPORT NO:-QMTL/R-85436

<b>CLIENT:- CRYSTALLINE ENVIRONMENTAL SERVICES</b> P.O.Box.131463, Abu Dhabi, U.A.E Ph:02-6508930/0557167017		
<b>ATTN:MR. JASMAN PINTO</b>		
Product: Water	Sampling Method: BS 7592	Sample ID: QMTL/S-85436
Reference: <b>HEALTHPOINT HOSPITAL</b>	Sampled By: QMTL Representative	Date Received: 19.12.2019
<b>Sample Description :</b> Dental Chair- 8 Sample Onsite Temp: 24.2°C Sample Receipt Temp: 17°C	Source: Dental Chair	Date(s) Tested: 19-29.12.2019
	Sample Container: 1 X 1L PB	Date Reported: 29.12.2019
	Sample Appearance: Clear	Date of Sampling/Time: 19.12.2019/12.35 PM
On-Site Treatment: NA	Tested by Emp#: 55 & 74	Location: AUH

**TEST RESULTS**

PARAMETERS	TEST METHODS	UNITS	LIMITS	RESULTS
<b>LEGIONELLA ANALYSIS</b>				
Legionella	CCFRA 3.9.1 (5 <sup>th</sup> Edition)	cfu/litre	<1	<b>Not Detected</b>
<b>IDENTIFICATION</b>				
Legionella Species	CCFRA 3.9.1 (5 <sup>th</sup> Edition)	-	-	<b>Absent</b>
Legionella pneumophila sero group 1	CCFRA 3.9.1 (5 <sup>th</sup> Edition)	-	-	<b>Absent</b>
Legionella pneumophila sero group 2-14	CCFRA 3.9.1 (5 <sup>th</sup> Edition)	-	-	<b>Absent</b>

REMARKS : The sample meets HAAD EHSMS Standard for Healthcare Providers- RSB Water Quality Regulations (4<sup>th</sup> Edition) January 2014 requirements for potable water with respect to the above tests only.

Key Words: CFU: Colony Forming Unit, BS: British Standards, CCFRA: Campden & Chorleywood Food Research Association, RSB: Regulation & Supervision Bureau, HAAD: The Health Authority- Abu Dhabi, EHSMS: Environment, Health and Safety Management System, NA: Not Applicable

Report prepared by Emp#: 75

Test Method deviation : None

The above test results are only applicable to the sample(s) referred above

**Reported By**

**ANILA GEORGE**  
**LABORATORY MANAGER**

The test Report shall not be reproduced (except in full) without the written approval of QMTL. When analysis is witnessed by us or carried out by sub contract labs, our responsibility is only to ensure that the analysis is conducted to standard test methods in accordance with industry accepted practice.

**END OF REPORT**

### TEST REPORT

REPORT NO:-QMTL/R-85417

<b>CLIENT:- CRYSTALLINE ENVIRONMENTAL SERVICES</b> P.O.Box.131463, Abu Dhabi, U.A.E Ph:02-6508930/0557167017		
<b>ATTN:MR. JASMAN PINTO</b>		
Product: Water	Sampling Method: APHA 9060 A	Sample ID: QMTL/S-85417
Reference: <b>HEALTHPOINT HOSPITAL</b>	Sampled By: QMTL Representative	Date Received: 19.12.2019
<b>Sample Description :</b> Dental Chair- 9 Sample Onsite Temp: 23.4°C Sample Receipt Temp: 4°C	Source: Dental Chair	Date(s) Tested: 19-21.12.2019
	Sample Container: 1 X 500ml PB	Date Reported: 23.12.2019
	Sample Appearance: Clear	Date of Sampling/Time: 19.12.2019/12.12 PM
On-Site Treatment: NA	Tested by Emp#: 55 & 74	Location: AUH

### TEST RESULTS

PARAMETERS	TEST METHODS (APHA 23 <sup>RD</sup> EDITION)	UNITS	SPECIFICATIONS	RESULTS
<b>MICROBIOLOGICAL ANALYSIS</b>				
Total Bacterial Count	APHA 9215 B	CFU/ml	10@37°C	<1
Total Coliforms	APHA 9222 B	CFU/100ml	Zero	<b>Zero</b>
Faecal coliform	APHA 9222 D	CFU/100ml	Zero	<b>Zero</b>
Pseudomonas aeruginosa	APHA 9213 E	CFU/100ml	Zero	<b>Zero</b>
Escherichia coli	APHA 9222 H & I	CFU/100ml	Zero	<b>Zero</b>

REMARKS : The sample meets the limits of HAAD EHSMS Standard for Healthcare Providers- RSB Water Quality Regulations (4<sup>th</sup> Edition) January 2014 requirements for potable water with respect to the above tests only.

Key Words:- CFU : Colony Forming Unit, APHA : American Public Health Association, RSB : Regulation & Supervision Bureau HAAD : The Health – Authority- Abu Dhabi, EHSMS : Environment, Health and Safety Management System, NA : Not Applicable

Report prepared by Emp#: 75

Test Method deviation: None

The above test results are only applicable to the sample(s) referred above

Reported By

**ANILA GEORGE**  
**LABORATORY MANAGER**

The test-Report shall not be reproduced (except in full) without the written approval of QMTL. When analysis is witnessed by us or carried out by sub contract labs, our responsibility is only to ensure that the analysis is conducted to standard test methods in accordance with industry accepted practice.

END OF REPORT

### TEST REPORT

REPORT NO:-QMTL/R-85418

<b>CLIENT:- CRYSTALLINE ENVIRONMENTAL SERVICES</b> P.O.Box.131463, Abu Dhabi, U.A.E Ph:02-6508930/0557167017		
<b>ATTN:MR. JASMAN PINTO</b>		
Product: Water	Sampling Method: APHA 9060 A	Sample ID: QMTL/S-85418
Reference: <b>HEALTHPOINT HOSPITAL</b>	Sampled By: QMTL Representative	Date Received: 19.12.2019
<b>Sample Description :</b> Dental Chair- 10 Sample Onsite Temp: 23.9°C Sample Receipt Temp: 4°C	Source: Dental Chair	Date(s) Tested: 19-21.12.2019
	Sample Container: 1 X 500ml PB	Date Reported: 23.12.2019
	Sample Appearance: Clear	Date of Sampling/Time: 19.12.2019/12.27 PM
On-Site Treatment: NA	Tested by Emp#: 55 & 74	Location: AUH

### TEST RESULTS

PARAMETERS	TEST METHODS (APHA 23 <sup>RD</sup> EDITION)	UNITS	SPECIFICATIONS	RESULTS
<b>MICROBIOLOGICAL ANALYSIS</b>				
Total Bacterial Count	APHA 9215 B	CFU/ml	10@37°C	<1
Total Coliforms	APHA 9222 B	CFU/100ml	Zero	<b>Zero</b>
Faecal coliform	APHA 9222 D	CFU/100ml	Zero	<b>Zero</b>
Pseudomonas aeruginosa	APHA 9213 E	CFU/100ml	Zero	<b>Zero</b>
Escherichia coli	APHA 9222 H & I	CFU/100ml	Zero	<b>Zero</b>

REMARKS : The sample meets the limits of HAAD EHSMS Standard for Healthcare Providers- RSB Water Quality Regulations (4<sup>th</sup> Edition) January 2014 requirements for potable water with respect to the above tests only.

Key Words:- CFU : Colony Forming Unit, APHA : American Public Health Association, RSB : Regulation & Supervision Bureau HAAD : The Health – Authority- Abu Dhabi, EHSMS : Environment, Health and Safety Management System, NA : Not Applicable

Report prepared by Emp#: 75

Test Method deviation: None

The above test results are only applicable to the sample(s) referred above

**Reported By**

**ANILA GEORGE**  
**LABORATORY MANAGER**

The test Report shall not be reproduced (except in full) without the written approval of QMTL. When analysis is witnessed by us or carried out by sub contract labs, our responsibility is only to ensure that the analysis is conducted to standard test methods in accordance with industry accepted practice.

END OF REPORT

### TEST REPORT

REPORT NO:-QMTL/R-85437

<b>CLIENT:- CRYSTALLINE ENVIRONMENTAL SERVICES</b>		
P.O.Box.131463, Abu Dhabi, U.A.E Ph:02-6508930/0557167017		
<b>ATTN:MR. JASMAN PINTO</b>		
Product: Water	Sampling Method: BS 7592	Sample ID: QMTL/S-85437
Reference: <b>HEALTHPOINT HOSPITAL</b>	Sampled By: QMTL Representative	Date Received: 19.12.2019
<b>Sample Description :</b> Dental Chair- 10A Sample Onsite Temp: 24.6°C Sample Receipt Temp: 17°C	Source: Dental Chair	Date(s) Tested: 19-29.12.2019
	Sample Container: 1 X 1L PB	Date Reported: 29.12.2019
	Sample Appearance: Clear	Date of Sampling/Time: 19.12.2019/12.38 PM
On-Site Treatment: NA	Tested by Emp#: 55 & 74	Location: AUH

### TEST RESULTS

PARAMETERS	TEST METHODS	UNITS	LIMITS	RESULTS
<b>LEGIONELLA ANALYSIS</b>				
Legionella	CCFRA 3.9.1 (5 <sup>th</sup> Edition)	cfu/litre	<1	<b>Not Detected</b>
<b>IDENTIFICATION</b>				
Legionella Species	CCFRA 3.9.1 (5 <sup>th</sup> Edition)	-	-	<b>Absent</b>
Legionella pneumophila sero group 1	CCFRA 3.9.1 (5 <sup>th</sup> Edition)	-	-	<b>Absent</b>
Legionella pneumophila sero group 2-14	CCFRA 3.9.1 (5 <sup>th</sup> Edition)	-	-	<b>Absent</b>

REMARKS : The sample meets HAAD EHSMS Standard for Healthcare Providers- RSB Water Quality Regulations (4<sup>th</sup> Edition) January 2014 requirements for potable water with respect to the above tests only.

Key Words: CFU: Colony Forming Unit, BS: British Standards, CCFRA: Campden & Chorleywood Food Research Association, RSB: Regulation & Supervision Bureau, HAAD: The Health Authority- Abu Dhabi, EHSMS: Environment, Health and Safety Management System, NA: Not Applicable

Report prepared by Emp#: 75

Test Method deviation : None

The above test results are only applicable to the sample(s) referred above

**Reported By**

**ANILA GEORGE**  
**LABORATORY MANAGER**

The test Report shall not be reproduced (except in full) without the written approval of QMTL. When analysis is witnessed by us or carried out by sub contract labs, our responsibility is only to ensure that the analysis is conducted to standard test methods in accordance with industry accepted practice.

END OF REPORT

**TEST REPORT**

REPORT NO:-QMTL/R-85419

<b>CLIENT:- CRYSTALLINE ENVIRONMENTAL SERVICES</b> P.O.Box.131463, Abu Dhabi, U.A.E Ph:02-6508930/0557167017		
<b>ATTN:MR. JASMAN PINTO</b>		
Product: Water	Sampling Method: APHA 9060 A	Sample ID: QMTL/S-85419
Reference: <b>HEALTHPOINT HOSPITAL</b>	Sampled By: QMTL Representative	Date Received: 19.12.2019
<b>Sample Description :</b> Dental Chair- 11 Sample Onsite Temp: 23°C Sample Receipt Temp: 4°C	Source: Dental Chair	Date(s) Tested: 19-21.12.2019
	Sample Container: 1 X 500ml PB	Date Reported: 23.12.2019
	Sample Appearance: Clear	Date of Sampling/Time: 19.12.2019/12.48 PM
On-Site Treatment: NA	Tested by Emp#: 55 & 74	Location: AUH

**TEST RESULTS**

PARAMETERS	TEST METHODS (APHA 23 <sup>RD</sup> EDITION)	UNITS	SPECIFICATIONS	RESULTS
<b>MICROBIOLOGICAL ANALYSIS</b>				
Total Bacterial Count	APHA 9215 B	CFU/ml	10@37°C	<1
Total Coliforms	APHA 9222 B	CFU/100ml	Zero	<b>Zero</b>
Faecal coliform	APHA 9222 D	CFU/100ml	Zero	<b>Zero</b>
Pseudomonas aeruginosa	APHA 9213 E	CFU/100ml	Zero	<b>Zero</b>
Escherichia coli	APHA 9222 H & I	CFU/100ml	Zero	<b>Zero</b>

REMARKS : The sample meets the limits of HAAD EHSMS Standard for Healthcare Providers- RSB Water Quality Regulations (4<sup>th</sup> Edition) January 2014 requirements for potable water with respect to the above tests only.

Key Words:- CFU : Colony Forming Unit, APHA : American Public Health Association, RSB : Regulation & Supervision Bureau HAAD : The Health – Authority- Abu Dhabi, EHSMS : Environment, Health and Safety Management System, NA : Not Applicable

Report prepared by Emp#: 75

Test Method deviation : None

The above test results are only applicable to the sample(s) referred above

**Reported By**

**ANILA GEORGE**  
**LABORATORY MANAGER**

The test-Report shall not be reproduced (except in full) without the written approval of QMTL. When analysis is witnessed by us or carried out by sub contract labs, our responsibility is only to ensure that the analysis is conducted to standard test methods in accordance with industry accepted practice.

END OF REPORT

**TEST REPORT**

REPORT NO:-QMTL/R-85420

<b>CLIENT:- CRYSTALLINE ENVIRONMENTAL SERVICES</b> P.O.Box.131463, Abu Dhabi, U.A.E Ph:02-6508930/0557167017		
<b>ATTN:MR. JASMAN PINTO</b>		
Product: Water	Sampling Method: APHA 9060 A	Sample ID: QMTL/S-85420
Reference: <b>HEALTHPOINT HOSPITAL</b>	Sampled By: QMTL Representative	Date Received: 19.12.2019
<b>Sample Description :</b> Dental Chair- 12 Sample Onsite Temp: 24°C Sample Receipt Temp: 4°C	Source: Dental Chair	Date(s) Tested: 19-21.12.2019
	Sample Container: 1 X 500ml PB	Date Reported: 23.12.2019
	Sample Appearance: Clear	Date of Sampling/Time: 19.12.2019/12.48 PM
On-Site Treatment: NA	Tested by Emp#: 55 & 74	Location: AUH

**TEST RESULTS**

PARAMETERS	TEST METHODS (APHA 23 <sup>RD</sup> EDITION)	UNITS	SPECIFICATIONS	RESULTS
<b>MICROBIOLOGICAL ANALYSIS</b>				
Total Bacterial Count	APHA 9215 B	CFU/ml	10@37°C	<1
Total Coliforms	APHA 9222 B	CFU/100ml	Zero	<b>Zero</b>
Faecal coliform	APHA 9222 D	CFU/100ml	Zero	<b>Zero</b>
Pseudomonas aeruginosa	APHA 9213 E	CFU/100ml	Zero	<b>Zero</b>
Escherichia coli	APHA 9222 H & I	CFU/100ml	Zero	<b>Zero</b>

REMARKS : The sample meets the limits of HAAD EHSMS Standard for Healthcare Providers- RSB Water Quality Regulations (4<sup>th</sup> Edition) January 2014 requirements for potable water with respect to the above tests only.

Key Words:- CFU : Colony Forming Unit, APHA : American Public Health Association, RSB : Regulation & Supervision Bureau HAAD : The Health – Authority- Abu Dhabi, EHSMS : Environment, Health and Safety Management System, NA : Not Applicable

Report prepared by Emp#: 75  
Test Method deviation : None  
The above test results are only applicable to the sample(s) referred above

**Reported By**

**ANILA GEORGE**  
**LABORATORY MANAGER**

The test Report shall not be reproduced (except in full) without the written approval of QMTL. When analysis is witnessed by us or carried out by sub contract labs, our responsibility is only to ensure that the analysis is conducted to standard test methods in accordance with industry accepted practice.

**END OF REPORT**